
















# EIGHTCHECK-3WP ASSAY SHEET














## Normal Level

### For XP series

**LOT** 4195 0822 

Parameters		Mean	Limit %	Limit #	Range	Barcode
WBC	[10 <sup>3</sup> /μL]	7.0	7.0	0.5	7.5	
					6.5	
RBC	[10 <sup>6</sup> /μL]	4.43	4.0	0.18	4.61	
					4.25	
HGB	[g/dL]	11.7	4.0	0.5	12.2	
					11.2	
	[mmol/L]	7.2	4.0	0.3	7.5	
					6.9	
HCT	[%]	33.3	7.5	2.5	35.8	
					30.8	
	[L/L]	0.333	7.5	0.025	0.358	
					0.308	
MCV	[fL]	75.2	5.0	3.8	79.0	
					71.4	
MCH	[pg]	26.4	5.5	1.5	27.9	
					24.9	
	[amol]	1639	5.5	90	1729	
					1549	
MCHC	[g/dL]	35.1	7.0	2.5	37.6	
					32.6	
	[mmol/L]	21.8	7.0	1.5	23.3	
					20.3	
PLT	[10 <sup>3</sup> /μL]	227	15	34	261	
					193	
W-SCR / LYM	[%]	30.6	15.0	4.6	35.2	
					26.0	

 19-Oct-2024 

Parameters		Mean	Limit %	Limit #	Range	Barcode
W-MCR / MXD	[%]	11.4	25.0	2.9	14.3	
					8.5	
W-LCR / NEUT <sup>1)</sup>	[%]	58.0	7.0	4.1	62.1	
					53.9	
W-SCC / LYM <sup>1)</sup>	[10 <sup>3</sup> /μL]	2.1	15.0	0.3	2.4	
					1.8	
W-MCC / MXD <sup>1)</sup>	[10 <sup>3</sup> /μL]	0.8	25.0	0.2	1.0	
					0.6	
W-LCC / NEUT <sup>1)</sup>	[10 <sup>3</sup> /μL]	4.1	10.0	0.4	4.5	
					3.7	
W-SMV	[fL]	59.3	18.0	10.7	70.0	
					48.6	
W-LMV	[fL]	190.4	15.0	28.6	219.0	
					161.8	
RDW-SD	[fL]	36.2	15.0	5.4	41.6	
					30.8	
RDW-CV	[%]	11.6	15.0	1.7	13.3	
					9.9	
PDW	[fL]	8.4	10.0	0.8	9.2	
					7.6	
MPV	[fL]	9.6	8.0	0.8	10.4	
					8.8	
P-LCR	[%]	16.9	60.0	10.1	27.0	
					6.8	
PCT	[%]	0.22	24.0	0.05	0.27	
					0.17	

1) These parameters values should be used as a reference only.