

# EIGHTCHECK-3WP ASSAY SHEET


## Normal Level


### For XP series



**LOT** 4279 0822  


 11-Jan-2025

Parameters	Mean	Limit %	Limit #	Range	
WBC	[10 <sup>3</sup> /μL]	7.1	7.0	0.5	7.6 6.6
					

Parameters	Mean	Limit %	Limit #	Range	
PDW	[fL]	8.8	10.0	0.9	9.7 7.9
					

RBC	[10 <sup>6</sup> /μL]	4.45	4.0	0.18	4.63 4.27
					

MPV	[fL]	10.0	8.0	0.8	10.8 9.2
					


MCV	[fL]	76.3	5.0	3.8	80.1 72.5
					


P-LCR	[%]	20.2	60.0	12.1	32.3 8.1
					

PLT	[10 <sup>3</sup> /μL]	219	15.0	33	252 186
					


PCT	[%]	0.22	24.0	0.05	0.27 0.17
					

W-SCR / LYM	[%]	31.1	15.0	4.7	35.8 26.4
					


Conventional Units					
Parameters	Mean	Limit %	Limit #	Range	
HGB	[g/dL]	11.9	4.0	0.5	12.4 11.4
					

W-MCR / MXD	[%]	8.7	25.0	2.2	10.9 6.5
					

HCT	[%]	34.0	7.5	2.6	36.6 31.4
					

W-LCR / NEUT <sup>1)</sup>	[%]	60.2	7.0	4.2	64.4 56.0
					


W-SCC / LYM <sup>1)</sup>	[10 <sup>3</sup> /μL]	2.2	15.0	0.3	2.5 1.9
					


MCH	[pg]	26.7	5.5	1.5	28.2 25.2
					

W-MCC / MXD <sup>1)</sup>	[10 <sup>3</sup> /μL]	0.6	25.0	0.2	0.8 0.4
					


MCHC	[g/dL]	35.0	7.0	2.5	37.5 32.5
					

W-LCC / NEUT <sup>1)</sup>	[10 <sup>3</sup> /μL]	4.3	10.0	0.4	4.7 3.9
					


SI Units					
Parameters	Mean	Limit %	Limit #	Range	
HGB	[mmol/L]	7.4	4.0	0.3	7.7 7.1
					

W-SMV	[fL]	59.2	18.0	10.7	69.9 48.5
					


W-LMV	[fL]	190.3	15.0	28.5	218.8 161.8
					

HCT	[L/L]	0.340	7.5	0.026	0.366 0.314
					

RDW-SD	[fL]	36.3	15.0	5.4	41.7 30.9
					

MCH	[amol]	1660	5.5	91	1751 1569
					

RDW-CV	[%]	11.3	15.0	1.7	13.0 9.6
					

MCHC	[mmol/L]	21.7	7.0	1.5	23.2 20.2
					

<sup>1)</sup> These parameters values should be used as a reference only.  
 Revised 08/2024

