
















EIGHTCHECK-3WP ASSAY SHEET














High Level

For XP series

LOT 4195 0823 

Parameters		Mean	Limit %	Limit #	Range	Barcode
WBC	[10 ³ /μL]	17.9	7.0	1.3	19.2	
					16.6	
RBC	[10 ⁶ /μL]	5.37	4.0	0.21	5.58	
					5.16	
HGB	[g/dL]	16.5	4.0	0.7	17.2	
					15.8	
	[mmol/L]	10.3	4.0	0.4	10.7	
					9.9	
HCT	[%]	46.0	7.5	3.5	49.5	
					42.5	
	[L/L]	0.460	7.5	0.035	0.495	
0.425						
MCV	[fL]	85.6	5.0	4.3	89.9	
					81.3	
MCH	[pg]	30.7	5.5	1.7	32.4	
					29.0	
	[amol]	1907	5.5	105	2012	
					1802	
MCHC	[g/dL]	35.9	7.0	2.5	38.4	
					33.4	
	[mmol/L]	22.3	7.0	1.6	23.9	
					20.7	
PLT	[10 ³ /μL]	535	13	70	605	
					465	
W-SCR / LYM	[%]	35.5	15.0	5.3	40.8	
					30.2	

 19-Oct-2024 

Parameters		Mean	Limit %	Limit #	Range	Barcode
W-MCR / MXD	[%]	15.9	25.0	4.0	19.9	
					11.9	
W-LCR / NEUT ¹⁾	[%]	48.6	7.0	3.4	52.0	
					45.2	
W-SCC / LYM ¹⁾	[10 ³ /μL]	6.4	15.0	1.0	7.4	
					5.4	
W-MCC / MXD ¹⁾	[10 ³ /μL]	2.8	25.0	0.7	3.5	
					2.1	
W-LCC / NEUT ¹⁾	[10 ³ /μL]	8.7	10.0	0.9	9.6	
					7.8	
W-SMV	[fL]	60.3	18.0	10.9	71.2	
					49.4	
W-LMV	[fL]	181.8	15.0	27.3	209.1	
					154.5	
RDW-SD	[fL]	35.8	15.0	5.4	41.2	
					30.4	
RDW-CV	[%]	9.8	15.0	1.5	11.3	
					8.3	
PDW	[fL]	8.5	10.0	0.9	9.4	
					7.6	
MPV	[fL]	9.6	8.0	0.8	10.4	
					8.8	
P-LCR	[%]	17.2	60.0	10.3	27.5	
					6.9	
PCT	[%]	0.51	20.0	0.10	0.61	
					0.41	

1) These parameters values should be used as a reference only.