
















# EIGHTCHECK-3WP ASSAY SHEET














## Low Level

### For XP series

**LOT** 4195 0821 

Parameters		Mean	Limit %	Limit #	Range	Barcode
WBC	[10 <sup>3</sup> /μL]	3.5	12.0	0.4	3.9	
					3.1	
RBC	[10 <sup>6</sup> /μL]	2.69	5.0	0.13	2.82	
					2.56	
HGB	[g/dL]	6.4	5.0	0.3	6.7	
					6.1	
	[mmol/L]	4.0	5.0	0.2	4.2	
					3.8	
HCT	[%]	18.8	8.5	1.6	20.4	
					17.2	
	[L/L]	0.188	8.5	0.016	0.204	
0.172						
MCV	[fL]	69.9	6.0	4.2	74.1	
					65.7	
MCH	[pg]	23.8	6.0	1.4	25.2	
					22.4	
	[amol]	1477	6.0	89	1566	
1388						
MCHC	[g/dL]	34.0	8.0	2.7	36.7	
					31.3	
	[mmol/L]	21.1	8.0	1.7	22.8	
19.4						
PLT	[10 <sup>3</sup> /μL]	85	30	26	111	
					59	
W-SCR / LYM	[%]	20.2	25.0	5.1	25.3	
					15.1	

 19-Oct-2024 

Parameters		Mean	Limit %	Limit #	Range	Barcode
W-MCR / MXD	[%]	10.2	30.0	3.1	13.3	
					7.1	
W-LCR / NEUT <sup>1)</sup>	[%]	69.6	15.0	10.4	80.0	
					59.2	
W-SCC / LYM <sup>1)</sup>	[10 <sup>3</sup> /μL]	0.7	25.0	0.2	0.9	
					0.5	
W-MCC / MXD <sup>1)</sup>	[10 <sup>3</sup> /μL]	0.4	30.0	0.1	0.5	
					0.3	
W-LCC / NEUT <sup>1)</sup>	[10 <sup>3</sup> /μL]	2.4	20.0	0.5	2.9	
					1.9	
W-SMV	[fL]	59.2	20.0	11.8	71.0	
					47.4	
W-LMV	[fL]	190.3	20.0	38.1	228.4	
					152.2	
RDW-SD	[fL]	38.5	20.0	7.7	46.2	
					30.8	
RDW-CV	[%]	14.0	20.0	2.8	16.8	
					11.2	
PDW	[fL]	8.3	20.0	1.7	10.0	
					6.6	
MPV	[fL]	9.5	10.0	1.0	10.5	
					8.5	
P-LCR	[%]	17.5	60.0	10.5	28.0	
					7.0	
PCT	[%]	0.08	43.0	0.03	0.11	
					0.05	

1) These parameters values should be used as a reference only.