
















# EIGHTCHECK-3WP ASSAY SHEET













## Normal Level

### For pocH-100i

**LOT** 4195 0822 

Parameters		Mean	Limit %	Limit #	Range	Barcode
WBC	[10 <sup>3</sup> /μL]	6.9	9.0	0.6	7.5	
					6.3	
RBC	[10 <sup>6</sup> /μL]	4.57	5.0	0.23	4.80	
					4.34	
HGB	[g/dL]	11.9	5.0	0.6	12.5	
					11.3	
	[mmol/L]	7.4	5.0	0.4	7.8	
					7.0	
HCT	[%]	35.6	8.0	2.8	38.4	
					32.8	
	[L/L]	0.356	8.0	0.028	0.384	
					0.328	
MCV	[fL]	78.0	6.0	4.7	82.7	
					73.3	
MCH	[pg]	26.0	6.0	1.6	27.6	
					24.4	
	[amol]	1616	6.0	97	1713	
					1519	
MCHC	[g/dL]	33.4	7.0	2.3	35.7	
					31.1	
	[mmol/L]	20.7	7.0	1.4	22.1	
					19.3	
PLT	[10 <sup>3</sup> /μL]	215	15	32	247	
					183	
W-SCR / LYM	[%]	31.2	15.0	4.7	35.9	
					26.5	

 19-Oct-2024 

Parameters		Mean	Limit %	Limit #	Range	Barcode
W-MCR / MXD	[%]	11.3	28.0	3.2	14.5	
					8.1	
W-LCR / NEUT <sup>1)</sup>	[%]	57.5	7.0	4.0	61.5	
					53.5	
W-SCC / LYM <sup>1)</sup>	[10 <sup>3</sup> /μL]	2.2	15.0	0.3	2.5	
					1.9	
W-MCC / MXD <sup>1)</sup>	[10 <sup>3</sup> /μL]	0.8	28.0	0.2	1.0	
					0.6	
W-LCC / NEUT <sup>1)</sup>	[10 <sup>3</sup> /μL]	3.9	12.0	0.5	4.4	
					3.4	
W-SMV	[fL]	64.0	18.0	11.5	75.5	
					52.5	
W-LMV	[fL]	202.1	15.0	30.3	232.4	
					171.8	
RDW-SD	[fL]	42.9	25.0	10.7	53.6	
					32.2	
RDW-CV	[%]	15.8	25.0	4.0	19.8	
					11.8	
PDW	[fL]	9.0	10.0	0.9	9.9	
					8.1	
MPV	[fL]	10.4	8.0	0.8	11.2	
					9.6	
P-LCR	[%]	19.5	60.0	11.7	31.2	
					7.8	

1) These parameters values should be used as a reference only.

