
















EIGHTCHECK-3WP ASSAY SHEET













Low Level

For pocH-100i

LOT 4195 0821 

| Parameters | | Mean | Limit % | Limit # | Range | Barcode |
|-------------|-----------------------|-------|---------|---------|-------|--------------------------------------------------------------------------------------|
| WBC | [10 ³ /μL] | 3.4 | 14.0 | 0.5 | 3.9 |  |
| | | | | | 2.9 | |
| RBC | [10 ⁶ /μL] | 2.74 | 6.0 | 0.16 | 2.90 |  |
| | | | | | 2.58 | |
| HGB | [g/dL] | 6.5 | 6.0 | 0.4 | 6.9 |  |
| | | | | | 6.1 | |
| | [mmol/L] | 4.0 | 6.0 | 0.2 | 4.2 |  |
| | | | | | 3.8 | |
| HCT | [%] | 19.9 | 8.5 | 1.7 | 21.6 |  |
| | | | | | 18.2 | |
| | [L/L] | 0.199 | 8.5 | 0.017 | 0.216 |  |
| 0.182 | | | | | | |
| MCV | [fL] | 72.5 | 7.0 | 5.1 | 77.6 |  |
| | | | | | 67.4 | |
| MCH | [pg] | 23.7 | 7.0 | 1.7 | 25.4 |  |
| | | | | | 22.0 | |
| | [amol] | 1472 | 7.0 | 103 | 1575 |  |
| 1369 | | | | | | |
| MCHC | [g/dL] | 32.7 | 8.0 | 2.6 | 35.3 |  |
| | | | | | 30.1 | |
| | [mmol/L] | 20.3 | 8.0 | 1.6 | 21.9 |  |
| | | | | | 18.7 | |
| PLT | [10 ³ /μL] | 81 | 30 | 24 | 105 |  |
| | | | | | 57 | |
| W-SCR / LYM | [%] | 20.4 | 25.0 | 5.1 | 25.5 |  |
| | | | | | 15.3 | |

 19-Oct-2024 

| Parameters | | Mean | Limit % | Limit # | Range | Barcode |
|----------------------------|-----------------------|-------|---------|---------|-------|---------------------------------------------------------------------------------------|
| W-MCR / MXD | [%] | 10.0 | 48.0 | 4.8 | 14.8 |  |
| | | | | | 5.2 | |
| W-LCR / NEUT ¹⁾ | [%] | 69.6 | 15.0 | 10.4 | 80.0 |  |
| | | | | | 59.2 | |
| W-SCC / LYM ¹⁾ | [10 ³ /μL] | 0.7 | 30.0 | 0.2 | 0.9 |  |
| | | | | | 0.5 | |
| W-MCC / MXD ¹⁾ | [10 ³ /μL] | 0.3 | 41.0 | 0.1 | 0.4 |  |
| | | | | | 0.2 | |
| W-LCC / NEUT ¹⁾ | [10 ³ /μL] | 2.4 | 20.0 | 0.5 | 2.9 |  |
| | | | | | 1.9 | |
| W-SMV | [fL] | 64.2 | 20.0 | 12.8 | 77.0 |  |
| | | | | | 51.4 | |
| W-LMV | [fL] | 202.0 | 20.0 | 40.4 | 242.4 |  |
| | | | | | 161.6 | |
| RDW-SD | [fL] | 44.3 | 30.0 | 13.3 | 57.6 |  |
| | | | | | 31.0 | |
| RDW-CV | [%] | 17.6 | 30.0 | 5.3 | 22.9 |  |
| | | | | | 12.3 | |
| PDW | [fL] | 8.8 | 20.0 | 1.8 | 10.6 |  |
| | | | | | 7.0 | |
| MPV | [fL] | 10.7 | 10.0 | 1.1 | 11.8 |  |
| | | | | | 9.6 | |
| P-LCR | [%] | 21.8 | 60.0 | 13.1 | 34.9 |  |
| | | | | | 8.7 | |

1) These parameters values should be used as a reference only.

