
















EIGHTCHECK-3WP ASSAY SHEET













High Level

For pocH-100i

LOT 4195 0823 

Parameters		Mean	Limit %	Limit #	Range	Barcode
WBC	[10 ³ /μL]	17.6	9.0	1.6	19.2	
					16.0	
RBC	[10 ⁶ /μL]	5.53	5.0	0.28	5.81	
					5.25	
HGB	[g/dL]	16.9	5.0	0.8	17.7	
					16.1	
	[mmol/L]	10.5	5.0	0.5	11.0	
					10.0	
HCT	[%]	49.4	8.0	4.0	53.4	
					45.4	
	[L/L]	0.494	8.0	0.040	0.534	
					0.454	
MCV	[fL]	89.3	6.0	5.4	94.7	
					83.9	
MCH	[pg]	30.6	5.5	1.7	32.3	
					28.9	
	[amol]	1897	5.5	104	2001	
					1793	
MCHC	[g/dL]	34.2	7.0	2.4	36.6	
					31.8	
	[mmol/L]	21.2	7.0	1.5	22.7	
					19.7	
PLT	[10 ³ /μL]	502	15	75	577	
					427	
W-SCR / LYM	[%]	38.0	15.0	5.7	43.7	
					32.3	

 19-Oct-2024 

Parameters		Mean	Limit %	Limit #	Range	Barcode
W-MCR / MXD	[%]	16.0	28.0	4.5	20.5	
					11.5	
W-LCR / NEUT ¹⁾	[%]	46.0	7.0	3.2	49.2	
					42.8	
W-SCC / LYM ¹⁾	[10 ³ /μL]	6.7	15.0	1.0	7.7	
					5.7	
W-MCC / MXD ¹⁾	[10 ³ /μL]	2.8	28.0	0.8	3.6	
					2.0	
W-LCC / NEUT ¹⁾	[10 ³ /μL]	8.1	12.0	1.0	9.1	
					7.1	
W-SMV	[fL]	64.4	18.0	11.6	76.0	
					52.8	
W-LMV	[fL]	191.2	15.0	28.7	219.9	
					162.5	
RDW-SD	[fL]	44.3	25.0	11.1	55.4	
					33.2	
RDW-CV	[%]	14.2	25.0	3.6	17.8	
					10.6	
PDW	[fL]	9.3	10.0	0.9	10.2	
					8.4	
MPV	[fL]	10.2	8.0	0.8	11.0	
					9.4	
P-LCR	[%]	18.7	60.0	11.2	29.9	
					7.5	

1) These parameters values should be used as a reference only.

